



## Declaration of Conformity

*According to Engineering Report # ER1190829-PDC*

**Manufacturer's Name:** IBA Dosimetry of America Inc. dba Radcal  
**Manufacturer's Address:** 426 West Duarte Road  
Monrovia, CA 91016  
U.S.A.

**Declares that the Product**  
**Product Name:** Patient Dose Calibrator  
**Model Number(s):** PDC-01  
**Product Accessories:** All

### Conforms to the following Product Specification:

#### EMC: \*

EN 61326-1 (2013) Group 1 Class B  
EN 61000-4-2  
EN 61000-4-3  
EN 61000-4-4  
EN 61000-4-6  
EN 61000-4-5  
EN 61000-4-11  
EN 61000-3-2  
EN 61000-3-3

#### PERFORMANCE:

IEC 61674

The product herewith complies with the RFI-emissions requirements and immunity requirements and carries the CE marking accordingly.

A handwritten signature in black ink, appearing to read "E. MacIntosh".

E. MacIntosh  
Technical Manager  
August 29, 2019

\* Tested with 3m USB cable